

04/15/15. Sample not preserved to pH <2 SU. BOD, carbonaceous: Result reported as Sampling Error (SE) 04/13/15. Sample flow weighted incorrectly. BOD, carbonaceous: Result reported as Lab Error (LE) 04/20/15. Bottle series duplicates failed precision.

Attachments

No attachments.

Report Last Saved By

AUSTIN, CITY OF (S AUSTIN REG)

User: jane.burazer@austintexas.gov
Name: Jane Burazer
E-Mail: jane.burazer@austintexas.gov

Date/Time: 2015-05-13 19:34 (Time Zone: -05:00)